

Occupational hygiene monitoring of phenol

Principle

The procedure utilizes sampling tubes containing XAD-7 adsorbent. XAD-7 is an organic polymer. When air is sucked through the sample tube using a calibrated precision air sampling pump, phenol is adsorbed to the surface. During analysis phenol is extracted from the adsorbent and analyzed using high performance liquid chromatography.

